

購買（英文）規格書

COMMODITY DESCRIPTION

關稅分類番號 H.S. NO.	品目番號 ITEM NO.	品目 및 規格 DESCRIPTION	單位 UNIT	數量 QUANTITY
		초음파세척기	Set	2

A. Features:

1. For internal and external cleaning of fine surgical instruments, da Vinci instruments and other lumen instruments
2. High powered instrument capacity sonic irrigator incorporating thermal disinfection, designed to be extremely versatile with specific
3. Increased cleaning effect by attaching heater
4. Provided at the base of the unit for opening the up-down load tray
5. Safety device is attached to operate ultrasonic when it has proper water level
6. The cleaner is durable because the whole body is constructed of stainless steel
7. Stage control&check is possible by control panel during the whole process.
8. Water supply, Drainage can be automatically operated by a motor valve
9. Controlling washing time, ultrasonic output, temperature, spray time and water level control is available.
10. Convenient for moving & fixing by locking caster wheels.
11. Sweep Function By changing the ultrasonic frequency continuously, efficient cleaning is available.

B. Specifications:

1. Capacity : 174L or more
2. Dimension (Inside W X D X H) : 1260mm X 864mm X 1,105mm or less
3. Ultrasonic Transducer : 38KHZ more
4. Ultrasonic Power : 2,000W (70L 1,200W) or more
5. Material : SUS 316(IN) SUS 304(OUT) or Available
6. Possible to open the lid automatically and manually because the lid is separated with the load tray
7. Door Open : Motorized
8. Power : 1Ph 220V or 3Ph 380V
9. Available Function : Air spray or none

C. Configurations : (A or B)

A

- | | |
|----------------------------|------|
| 1. Main Unit(104L and 70L) | 2 EA |
| 2. Standard Tray | 3 EA |
| 3. Robot Tray | 2 EA |
| 4. O/S Manual | 1 EA |

B

- | | |
|--------------------|------|
| 1. Main Unit | 1 EA |
| 2. Standard Basket | 1 EA |
| 3. Robot Tray | 1 EA |
| 4. O/S Manual | 1 EA |

D. Remarks :

1. The equipment should be installation by contractor.
2. Five years warranty should be provided after the performance test.
3. Manufactured products must be delivered within one year based on the inspection date.
4. Products that are not scheduled to be discontinued within one year after delivery must be delivered
5. Replace equipment with new equipment if the equipment fails three times within one year of installation
6. A report must be summited after conducting a preventive inspection in accordance with the hospital medical equipment inspection regulations within the warranty period.
7. The ultrasonic cleaner must be a product that has been verified to be able to ultrasonically clean Davinch surgical roboot instruments.