

購 買 (英 文) 規 格 書

COMMODITY DESCRIPTION

關 稅 分 類 番 號 H.S. NO.	品 目 番 號 ITEM NO.	品 目 및 規 格 DESCRIPTION	單 位 UNIT	數 量 QUANTITY
		체외충격파쇄석기(ESWL)	set	1

A. Features:

1. Electro-Magnetic shockwave source
2. Multi-function lithotripsy renal stone, Ureteral stone, Gallbladder stone, Salivary stone, etc.
3. High-resolution Image FPD (Flat Panel Detector) system
4. Much less pain for patients and much satisfaction from patients
5. Shock wave device linked to stone movement by breathing

B. Specifications:

1. Shockwave Generator
 - Shockwave source: Electro-Magnetic
 - Focal Depth: 160mm
 - Focal Size: 1.9mm X 25.5mm or more
 - Focal Pressure: Max. 110MPa or more
 - Energy level: 10 steps
 - Shockwave Interval: 0.6sec ~ 2sec or more
2. X-ray system
 - Fluoroscopy kVp range: 40 ~ 110kVp or more
 - Fluoroscopy mA range: 0.5 ~ 6.0mA or more
 - Flat panel detector size: 9" or more
 - Focal spot: 0.5mm / 1.5mm or more
 - C-arm movement range: $\pm 30^{\circ}$
 - Fluoroscopy mode: Normal/Plus/Boost
3. Treatment table
 - Motorized movement
 - Length: 2000mm or more
 - Width: 680mm or more
 - Vertical movement: 200mm or more
 - Transversal movement: 150mm

- Longitudinal movement: 150mm
 - Driving method: Remote control
4. OP console
- Monitor: 24" or more
 - RAM: 16GB
 - O/S: Windows 10
 - Control Method: PC and Remote

C. Configurations:

1. Electro-magnetic source shockwave system	1set
2. Shockwave invert type generator system	1set
3. High Frequency invert X-ray system	1set
4. Treatment Table	1set
5. Water system	1set
6. Remote control	1set
7. Main console PC system	1set
8. User program	1set
9. Operating / Service manual	each 1bk
10. Water membrane	1ea

D. Remarks:

1. The equipment should be installation by contractor.
2. Three years warranty after the installation and correct operation.
3. Manufactured products must be delivered within one year based on the inspection date.
4. Products that are not scheduled to be discontinued within one year after delivery must be delivered.
5. A report must be submitted after conducting a preventive inspection in accordance with the hospital medical equipment inspection regulations within the warranty period.
6. It includes radiation protection facilities, wallpaper, and floorboard interior work in the extracorporeal shock wave testing room.
7. 보증기간 제외품목: 없음