

購買（英文）規格書

COMMODITY DESCRIPTION

關稅分類番 號 H.S. NO.	品目番號 ITEM NO.	品 目 및 規 格 DESCRIPTION	單 位 UNIT	數 量 QUANTITY
		플라즈마멸균기	SET	1

A. Features

1. Low Temperature Plasma Sterilization System
2. Sterilant Cassette Supply System
3. Automatic Sterilant Discharge System
4. Equipped with Four Sterilization Modes
5. Sterilization Cycle Time: Up to 62 Minutes
6. Equipped with an Automatic Drying System
7. Chamber Sterilant Concentration Monitoring System
8. Sliding Door Type
9. Compatible with da Vinci 5 Endoscopes for Sterilization
10. Certifications: KFDA, PMDA/MHLW, CFDA, CE (MDR), GMP or FDA

B. Specifications:

1. Dimensions (H × W × D): Within 1,800 × 830 × 1,100 mm
2. Weight (kg) : Within 440 kg

3. Equipped with Four Sterilization Modes
4. Sterilization Cycle Time: 19 ~ 62 Minutes
5. Chamber Sterilant Concentration Monitoring System
6. Sterilization Chamber
 - 1) Chamber Shape: Rectangular
 - 2) Dimension (H x W x D): Within 410 x 510 x 735mm
 - 3) Total Volume: Within 152L
 - 4) Usable Chamber Volume: Within 93.4L or more
7. Sterilant Type: Cassette (Single-Use or 5 Cycle)
8. Sterilant Concentration: Within 50%-59.6%
9. Sterilant Injection Method: Semi-Automatic or Manual
10. Automatic Sterilant Discharge System
11. Equipped with an Automatic Drying System
12. Door Type: Sliding Door
13. Door Control: Touchscreen or Foot Switch
14. Sterilization Indicators
 - 1) Biological indicators: BI Readout Within 30 Minutes
 - 2) Chemical indicators: Immediate Tape & Strip Verification
15. Display : Touchscreen, 10.2 Inches or Larger
16. Data Storage System: USB
17. Sterilization Completion Alarm: Melody
18. Automatic PM Kit Replacement Alert
19. Printer Outputs: Thermal Printer
20. Equipment Operation Monitoring System (Optional)

C. Configurations: (A or B)

A

1. Main unit	1 Unit
2. Sterilization agent (15ea / box)	2 Box
3. Chemical Indicator Strip (250ea / pack)	2 Pack
4. Chemical indicator tape (2rolls / pack)	1 Pack
5. Biological Indicator(30ea / box)	1 Box
6. Pouch Roll 150 (150mm × 70m)	2 Roll
7. Pouch Roll 250 (250mm × 70m)	2 Roll
8. Tray (2p/set)	1 Set
9. Incubator	1 Ea
10. Printer paper (5rolls / pack)	1 Pack
11. Protective gloves	1 Ea
12. User Manual	1 Ea
13. Service Manual	1 Ea

B

1. Main unit	1 Unit
2. Express Cycle	1 Ea
3. Operation & Service Manual	Each 1 Ea

D. Remarks:

1. The equipment should be installation by contractor.

2. Five years warranty should be provided after the performance test.
3. Manufactured products must be delivered within one year based on the inspection date.
4. Products that are not scheduled to be discontinued within one year after delivery must be delivered.
5. A report must be submitted after conducting a preventive inspection in accordance with the hospital medical equipment inspection regulations within the warranty period.
6. 무상보증제외품목 : 소독제, 테이프, 프린트 용지 등 소모품