

購 買 (英文) 規 格 書

COMMODITY DESCRIPTION

關 稅 分 類 番 號 H.S. NO.	品目番號 ITEM NO.	品 目 및 規 格 DESCRIPTION	單 位 UNIT	數 量 QUANTITY
		건조기(Drying Cabinet)	SET	1

A.Feature :

1. The drying cabinet is designed especially for drying of surgical instruments, anesthetic hoses & equipment, as well as other reusable items used in patient health care.
2. The sound and heat insulated double walls of the cabinet are made of ANSI 304 stainless steel. The cabinet is available as a single,double door model.
3. It secures the appropriate conditions for drying of cleaned and disinfected items by using HEPA filtered air at controlled temperature for various articles and materials.
4. The microprocessor controls the temperature process and the time of all cycles. It indicates visually and acoustically once the cycle finishes.
5. Heat exchanger for saving electricity and decreasing humidity exhaust.

B. Specification :

1. Dimension : 1900(H)*710(W)*785(D) or 1840(H)*850(W)*675(D)
2. Weight : 195kg or 200kg
3. Chamber size : 1400(H)*600(W)*600(D) or 1600(H)*640(W)*600(D)
4. Chamber capacity : 504L or 520L
5. Adjustment of the vertical distance between retractable shelves
6. Time adjustment range : 1-900minutes or more
7. Temperature : min. 45℃ ~ max. 80℃ or more
8. Hepa filter condition indication and double overheating protection
9. Filter traps 99.95% of all particles or more
10. Power : Single phase 220V 3.6KW or 3Φ4W 380V 3.5KW
11. Sound & heat insulated
12. Air flow rate 110m³/h or more

C. Configurations :

- | | |
|-------------------------------------|-----------|
| 1. Main Unit (with Standard Acc' y) | 1 SET |
| 2. HEPA Filter System(equipped) | 1 SET |
| 3. Operation & Service Manual | each 1 EA |

D. Remarks :

1. The equipment should be installation by contractor.
2. Three years warranty should be provided after the performance test.
3. Manufactured products must be delivered within one year based on the inspection date.
4. Products that are not scheduled to be discontinued within one

years after delivery must be delivered.

5. A report must be submitted after conducting a preventive inspection in According with the hospital medical equipment inspection regulations within the warranty period.

6. 보증기간내 소모품(HAPA Filter)은 보증제외