

購 買 (英 文) 規 格 書

COMMODITY DESCRIPTION

關 稅 分 類 番 號 H.S. NO.	品目番號 ITEM NO.	品 目 및 規 格 DESCRIPTION	單 位 UNIT	數 量 QUANTITY
		고압산소치료기	SET	1

A.Feature :

1. A Mono place chamber designed for treating patients with infectious risks or those who are conscious during treatment.
2. Designed to safety and effectively provide hyperbaric oxygen therapy above 3 ATA. All components comply with medical device certification standards and are constructed in a validated cylindrical form that can safely withstand high-pressure environments.
3. The chamber is a partially transparent cylindrical type, capable of pressurization over 3 ATA, helping to reduce claustrophobia for conscious patients and thereby ensuring treatment stability.
4. Designed for clinical convenience, allowing medical staff to visually monitor the patient and control pressure simultaneously.
5. Manufactured by a proven supplier whose equipment is widely used in top-tier general and university hospital, with validated stability and effectiveness. Stable product supply and management are also ensured.

6. Since immediate professional intervention is essential in emergencies, this is the only manufacturer that provides emergency A/S(after-sales service) to restore the chamber to normal operation upon fault detection.
7. Lightweight model suitable for installation in compact spaces, with a special material entrance to reduce feelings of enclosure and a user friendly control panel for intuitive operation.
8. Communication with the patient inside the chamber is possible via intercom.
9. Complies with ASME(American Society of Mechanical Engineers)

B. Specification :

1. Chamber Unit
 - (1)Max Internal Pressure : 2bar(3ATA)
 - (2)Weight : 550kg or less
 - (3)Dimensions(WxWxD)mm : 2,094 x 980 x 1392 or less
2. Entertainment Monitor : 32 inch
3. Cooling and Heating Unit : Available or No
4. Compressor Pressure : 6-8kgf/cm² or NO
5. Fire Extinguisher or Emergency Decompression
6. Power source : 1ph 220V

C. Configurations : (A or B)

A

- | | |
|----------------------------------|--------|
| 1. Chamber Body(Pressure Vessel) | 1 UNIT |
| 2. Control Console | 1 EA |

3. Entertainment Monitor	1 EA
4. Heating and Cooling System	1 EA
5. Compressor	1 EA
6. Fire Extinguisher	1 EA
7. Quick Manual	1 EA
8. Operation & Service Manual	each 1 EA
9. BIBS MASK	1 EA
10. Oxygen Durtain MASK	1 EA
11. Goggle-type MASK	1 EA

B

1. Monoplace Hyperbaric Oxygen Therapy Chamber	1 UNIT
2. Patient Mattress	1 SET
3. Patient Pillow	1 SET
4. Standard Accessories	1 SET
5. Quick Manual	1 EA
7. Operation & Service Manual	each 1 EA

D. Remarks :

1. The equipment should be installation by contractor.
2. Three years warranty should be provided after the performance test.
3. Manufactured products must be delivered within one year based on the inspection date.

4. Products that are not scheduled to be discontinued within one years after delivery must be delivered.
5. A report must be submitted after conducting a preventive inspection in According with the hospital medical equipment inspection regulations within the warranty period.
6. 구성품 중 환자 적용하는 MASK, Matteree, Pillow 는 보증기간에서 제외.