

購 買 (英 文) 規 格 書

COMMODITY DESCRIPTION

關 稅 分 類 番 號 H.S. NO.	品目番號 ITEM NO.	品 目 및 規 格 DESCRIPTION	單 位 UNIT	數 量 QUANTITY
		검진대	set	1

A. Features:

1. The low seat initial height allows easy access for patient as sitting on a chair.
2. The table is equipped with a preset and an auto-return device for speedy and accurate patient positioning.
3. Selective leg opening system allows patient to easily open and close right/left individually.
4. The emergency positioning is also provided.

B. Specifications:

1. Main Unit

- 1) Dimensions: 690(W)×1230(D)mm or more
- 2) Initial Seat Height: 450mm or less
- 3) Examination Seat Height: Max 1030mm or more
- 4) Preset: Backrest tilt up & Table up
- 5) Auto-return: Backrest tilt down & Table down
- 6) Drive Mechanism: Hydraulic or Motorized
- 7) Backrest Tilting: 70 ° up ~ -6 ° down or more
- 8) Control: Wireless foot control or Hand control
- 9) Input Voltage: 120 ~ 230VAC, 50/60Hz
- 10) Output Voltage: 6VDC or more
- 11) Patient Weight (Max): 135kg or more

12)Net Weight: 80.5kg or more

**C. Configurations: <A> or **

<A>

1. Main Unit	1unit
2. Wireless Foot Controller	1ea
3. Douche Tray	1ea
4. Leg Rest Cover	1box(100pcs)
5. Operating & Service Manual	each 1ea

1. Main Unit	1unit
2. Fold Up Foot Rest	1pair
3. Heavy Duty Leg Rests	1pair
4. Leg Rest Holder (Single)	2ea
5. Patient Arm Rest (Single)	2ea
6. Seat Cushion 170mm Cut-out	1ea
7. Two Motor Hand Control	1ea
8. Operating & Service Manual	each 1ea

D. Remarks:

1. The equipment should be installation by contractor.
2. Three years warranty after the installation and correct operation.
3. Manufactured products must be delivered within one year based on the inspection date.
4. Products that are not scheduled to be discontinued within one year after delivery must be delivered.
5. A report must be summited after conducting a preventive inspection in accordance with the hospital medical equipment inspection

- regulations within the warranty period.
6. 보증기간 제외품목: 없음